

广州市保为康劳保用品有限公司

Guangzhou Powecom Labor Insurance Supplies Co., Ltd



保为康<sup>®</sup>  
POWECOM

Guangzhou Powecom Labor Insurance Supplies Co., Ltd

Add: Floor 1, 3 and 4, Building 1, No. 43, Tuanjie Road, Xinya Street, Huadu District,  
Guangzhou City, Guangdong Province, P. R. China

Tel: +86-20-86600663



保为康®  
POWECOM

Description	KN95 Protective Mask
Item	KN95
Size	16*10cm
Package	10pcs*100bags
Pcs/Ctn	1000
Carton Size	68*36*36.5cm
Cbm/Ctn	0.0931cbm/ctn
Standard	GB2626 - 2006
Application	It is suitable for protection against harmful particles such as dust, PM2.5 smog particles, influenza and droplet, etc.









# Product



# Packging



# Carton

	<b>KN95</b>		<b>KN95</b>
 <b>KN95 PROTECTIVE MASK</b> GB2626-2006  1000pcs	Product Name: KN95 PROTECTIVE MASK Model No: KN95(Earloop style) Packaging: 1000pcs(10 x 100Bags) Carton Size: 68 x 36 x 36.5 cm Made In China	 <b>KN95 PROTECTIVE MASK</b> GB2626-2006  1000pcs	MANUFACTURER: Guangzhou Powecom Labor Insurance Supplies Co., Ltd. Add: floor 1, 3 and 4, Building 1, No. 43, Tuanjie Road, Xinya Street, Haodu District, Guangzhou City, Guangdong Province, P.R. China Tel: +86-20-86600663 Fax: +86-20-36092323 Made in China    

## Certificate





<b>合格证</b> Certificate	
品牌: 保为康 Brand: POWECOM	型号: KN95 Model No.: KN95
品名: KN95防护口罩 Item: KN95 Protective Mask	
数量: 10只/袋 Packaging: 10pcs/bag	等级: KN95 Grade: KN95
执行标准: GB2626-2006 KN95 Standard: GB2626-2006 KN95	
产品构成: 46% PP无纺布, 28%熔喷布, 26%热风棉 Composition: 46% PP Non-woven fabric, 28% melt blown fabric, 26% ES hot air cotton	
检验员: Inspector:	
生产日期: 2020/05/02 Date of Manufacture: 02/05/2020	
生产批次: 20200502 Batch No.: 20200502	
有效期: 3年 Validity: 3 years	
生产商: 广州市保为康劳保用品有限公司 Manufacturer: Guangzhou Powecom Labor Insurance Supplies Co., Ltd	
地址: 广东省广州市花都区新雅街团结路自编43号 1栋一、三、四层 Add: Floor 1, 3 and 4, Building 1, No. 43, Tuanjie Road, Xinya Street, Huadu District, Guangzhou City, Guangdong Province, P. R. China	

48 MM

70 MM

## Product instruction

74MM

<b>PRODUCT INSTRUCTION</b>	
<b>MODEL:</b> KN95 (Earloop style)	
<b>ITEM:</b> KN95 Protective Mask	
<b>APPLICATION:</b> Suitable for protection against powder, PM2.5 smog, particulates, saliva, dust etc. The face mask is good for wind proof and protection the face against cold weather.	
<b>COMPOSITION:</b> Outer layer PP Non-woven fabric Middle layer ES hot air cotton Non-woven fabric and efficiency electrostatic Melt-blown fabric Inner layer PP Non-woven fabric Executive standard GB 2626-2006 KN95 Non-oily particulate filtering efficiency ≥95%	
<b>STORAGE TEMPERATURE:</b> Temperature -10° C - 40° C, humidity < 80% Store in dry and cool place	
<b>VALIDITY:</b> Three years	
<b>INSTRUCTION FOR USE:</b>	
	
	
<b>CAUTIONS:</b>	
1. This product cannot be used to prevent harmful gases.	
2. Discard the face mask when it gets quite dirty or breathing resistance increase remarkably.	
3. Please avoid storing this product in the environment of high temperature, humidity or harmful chemicals.	
4. Preschoolers or patients with respiratory abnormalities can't use it.	
5. Everytime you use it, please check the wearing "adhesive" to ensure that the mask is worn correctly contamination.	
6. Not suitable for use in environments with toxic gases or oxygen concentrations below 19.5%.	
7. Cannot be used in microwave oven.	
<b>CUSTOMER SERVICE HOTLINE:</b> 400-008-5933	
<b>MANUFACTURER:</b> Guangzhou Powecom Labor Insurance Supplies Co., Ltd Add: Floor 1, 3 and 4, Building 1, No. 43, Tuanjie Road, Xinya Street, Huadu District, Guangzhou City, Guangdong Province, P. R. China	

210MM



# Test report

广州纤维产品检测研究院  
Guangzhou Fiber Product Testing and Research Institute

MA A  
CERTIFIED

## TEST REPORT

VERIFICATION WEBSITE: [www.gtt.net.cn](http://www.gtt.net.cn)  
VERIFICATION CODE: YPXT-2528-04

No: 200000861  
DATE: 2020-04-29

APPLICANT: Guangzhou Powecom Labor Insurance Supplies Co., Ltd.  
Floor 1, 3 and 4, Building 1, No. 43, Tuanjie Road, Xinya Street, Huadu District,  
Guangzhou City, Guangdong Province, P. R. China

Twenty-five (25) Pieces Of Powecom KN95 Mask  
STYLE NO.: 20200402

DATE RECEIVED/DATE TEST STARTED: 2020-04-29 16:59

CONCLUSION:

FILTRATION EFFICIENCY TO NaCl PARTICULATE MATTER	M
INSPIRATORY RESISTANCE	M
EXPIRATORY RESISTANCE	M

NOTE: "M" -MEET CLIENT'S REQUIREMENT "F" -BELOW CLIENT'S REQUIREMENT  
"---" -NO COMMENT "X" -HALF GRADE LOW "L" -ONE GRADE LOW

REMARK:  
IN THIS REPORT, THE SAMPLES ARE TESTED UNDER PRESCRIBED ENVIRONMENTAL CONDITIONS IN THE CORRESPONDING STANDARDS,  
UNLESS OTHERWISE NOTED.  
COPIES OF THE REPORT ARE VALID ONLY RE-STAMPED.  
WHEN THE INFORMATION PROVIDED BY THE CUSTOMER MAY AFFECT THE EFFECTIVENESS OF THE RESULTS, OUR INSTITUTE WILL  
NOT ASSUME ANY RESPONSIBILITY.

广州纤维产品检测研究院  
检验检测专用章

AUTHORIZED SIGNATORY: *张建功*  
HaoYang Zhang SENIOR ENGINEER

page 1 of 3


地址: 广州市海珠区滨江中路翠苑园35号之二 电话: 020-34402327 传真: 020-34402321

广州纤维产品检测研究院  
Guangzhou Fiber Product Testing and Research Institute

MA A  
CERTIFIED

## TEST REPORT

No: 200000861



广州纤维产品检测研究院  
检验检测专用章

page 2 of 3

地址: 广州市海珠区滨江中路翠苑园35号之二 电话: 020-34402327 传真: 020-34402321



扫描全能王 创建



扫描全能王 创建

# Test report

广州纤维产品检测研究院  
Guangzhou Fiber Product Testing and Research Institute

MA A  
QUALITY SYSTEM CERTIFIED

## TEST REPORT

No: 200000861

---

FILTRATION EFFICIENCY TO NaCl PARTICULATE MATTER (%)  
(GB 2626-2006 6.3 AIR FLOW: 95L/min AEROSOL: NaCl AEROSOL CONCENTRATION: 15mg/m<sup>3</sup>  
TEMPERATURE: 25.1°C RH: 38.2%)

UNTREATED SAMPLE:	REQUIREMENT
1# 99.09 2# 99.26	≥96.0
3# 99.18 4# 98.97	(KN95)
5# 99.01 6# 98.99	(GB 2626-2006)
7# 99.21 8# 99.14	
9# 98.79 10# 99.21	

---

INSPIRATORY RESISTANCE (Pa)  
(GB 2626-2006 6.5 HEAD MODEL: MEDIUM)

UNTREATED SAMPLE	REQUIREMENT
1#: 106.7	≤350
2#: 115.9	(GB 2626-2006)

REMARK: THE ITEM WAS CARRIED OUT AT No. 1-2, ZHUJIANG ROAD, CHAOTIAN INDUSTRIAL ZONE, SHILOU, PANYU DISTRICT, GUANGZHOU.

---

EXPIRATORY RESISTANCE (Pa)  
(GB 2626-2006 6.6 HEAD MODEL: MEDIUM)

UNTREATED SAMPLE	REQUIREMENT
1#: 88.8	≤250
2#: 87.5	(GB 2626-2006)

REMARK: THE ITEM WAS CARRIED OUT AT No. 1-2, ZHUJIANG ROAD, CHAOTIAN INDUSTRIAL ZONE, SHILOU, PANYU DISTRICT, GUANGZHOU.

End of Report

page 3 of 3

地址: 广州市海珠区滨江中路翠苑35号之二 电话: 020-34402327 传真: 020-34402321



扫描全能王 创建